

Excess Property for Salvage

The following items require EHS attention prior to transportation to salvage. If further information is required, please contact your Division Liaison.

Items	Examples	Action
Asbestos-containing items	Autoclaves, furnaces, incubators & ovens	Contact IH Must be in good condition & can be handled without producing airborne fibers. Asbestos must be abated if item will be scrapped
Bio-safety cabinets		Contact IH Decontamination is required
Chemical storage cabinets	Flammable storage lockers, acid and corrosive storage cabinets	Clean and remove visible residue. Contact IH Must be inspected
Electrical components	Capacitors & transformers	Electric shop drains oil. PCBs are a concern. Contact the Facilities Work Request Center -- xt.6274 http://iss15501.lbl.gov:591/wrc/
Fume hoods		Clean and remove visible residue. Contact IH Must be inspected
Gas cabinets		Clean and remove visible residue. Contact IH Must be inspected
Gas chromatographs	Electron capture detectors have sealed radiological sources	Contact RCT
Glove boxes		Clean and remove visible residue. Contact IH Must be inspected
Mercury-containing items	Barometers, coulter counters, manometers, switches & thermometers	Contact IH May need to drain mercury and/or dispose via Waste Management
Pumps		Pump shop must drain oil. Contact the Facilities Work Request Center (Ext. 6274) http://iss15501.lbl.gov:591/wrc/
Refrigerators & freezers	Those that have been used for chemical storage	Clean and remove visible residue. Contact IH Must be inspected
Shop equipment	Lathes & mills	Drain all fluids. Contact IH if equipment had contact with beryllium, lead or arsenic compounds. Contact RTC if there is possible radiological contamination
Static eliminators	Contain sealed radiological sources	Contact RCT
Syringes		Dispose in sharps container
Video monitors -CRTs, oscilloscopes	Non-functioning monitors that can't be repaired	Contact Generator Assistant and dispose of as waste
Washing equipment	Lab glassware washers, cage washers, etc. -- may contain corrosive cleaning fluids	Contact IH Must be inspected
X-ray-producing equipment		Contact RCT

NOTE: Costs for sample analyses and decontamination may be borne by the custodian